

ANALYSES BULLETIN	
SAMPLE	Veil tampon with applicator
Brand name	V-veil'UP
Manufactured by	V-VEIL-SHOP
Batch number	3 x (0916263)
BarCode	3 760 240 250 104
Packaging	Cardboardbox
Price	/
Quantity	3 x 6
Packaging Claims	0 - 6 g

<b>TAMPON</b>
---------------

DETERMINATIONS	ANALYSES		RESULTS	RECOMMENDATIONS POA NOT3 - Rév.3	X = Non-Compliant	
CHARACTERISTICS OF TAMPON	Net weight	g	0.5			
	Dimensions and Weight	Length	mm	/		
PERFORMANCE	Syngina test	Absorbed volume (average)	g	1.6	<6*	
		coefficient	g/g	2.9		
		Absorbed volume Tampon 1	g	1.5		
		Absorbed volume Tampon 2	g	1.6		
		Absorbed volume Tampon 3	g	1.6		
		Absorbed volume Tampon 4	g	1.6		
		Absorbed volume Tampon 5	g	1.6		
		Absorbed volume Tampon 6	g	1.5		
		Absorbed volume Tampon 7	g	1.5		
		Absorbed volume Tampon 8	g	1.6		
	Absorbed volume Tampon 9	g	1.5			
	Absorbed volume Tampon 10	g	1.5			
	Cord solidity	strength at break (average)	N	64.0	>30	
		strength at break cord 1	N	46.5		
strength at break cord 2		N	65.4			
strength at break cord 3		N	66.8			
strength at break cord 4		N	60.6			
	strength at break cord 5	N	80.5			

<b>CONCLUSION:</b>	Finishing	/
POA NOT3 - Rév.3	Absorption Syngina EDANA method	Conformable
	Other remark	/

\* absorption labelling

**In comparison with our recommendations : The product complies.**

Number of observations : 0.

Number of deviations : 0.

**Franck Facompré**

